SunLong Biotech Co.,LTD
Tel: 0086-571-56623320 Fax: 0086-571-56623318
E-mail: sales@sunlongbiotech.com
www.sunlongbiotech.com

Sunlong Medical™ 19-methyltestosterone/19-NT ELISA Kit

Catalog Number: ADA0404

Size: 15

Storage: 2 - 8 °C

Validity Period: 6 months

Sensitivity: 0.5 ng/mL

Assay range: 1.56 - 100 ng/mL

Used for the quantitative determination in plasma, serum, urine.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

SPECIFICATIONS

Catalog No.	ADA0404
Alternative Names	NADROLONE;17-beta-hydroxyoestr-4-en-3-one;17-HYDROXY-13-METHYL-2,6,7,8,9,10,11,12,14,15,16,17-DODECAHYDRO-1H-CYCLOPENTA[A]PHENANTHREN-3-ONE,5-ANDROSTENEDIONE;17b-hydroxyestra-4-en-3-one;17BETA-HYDROXY-4-ESTREN-3-ONE;17BETA-HYDROXY-19-NORANDROST-4-EN-3-ONE;19-NORTESTOSTERONE;19-NORANDROSTENOLONE
Detection method	Colorimetric
Sample type	Plasma, Serum, Urine
Assay type	Quantitative (Competitive)
Range	1.56 - 100 ng/mL
Sensitivity	0.5 ng/mL
Precision	CV
Recovery	80-120%
Shipping	2-8 °C
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition. Note