SunLong Biotech Co.,LTD
Tel: 0086-571-56623320 Fax: 0086-571-56623318
E-mail: sales@sunlongbiotech.com
www.sunlongbiotech.com

Sunlong Medical™ Medroxyprogesterone/MP ELISA Kit

Catalog Number: ADA0408

Size: 16

Storage: 2 - 8 °C

Validity Period: 6 months

Sensitivity: 0.5 ng/mL

Assay range: 1.56 - 100 ng/mL

Used for the quantitative determination in plasma, serum, urine.

FOR RESEARCH USE ONLY. NOT FOR USE IN DIAGNOSTIC PROCEDURES.

SPECIFICATIONS

Catalog No.	ADA0408
Alternative Names	PREGN-4-ENE-3,20-DIONE,17-HYDROXY-6-ALPHA-METHYL-;17··Hydroxy-6··methyl-4-pregnene-3,20-dione,17··Hydroxy-6··methylprogesterone,6··Methyl-5-pregnen-17··ol-3,20-dione,6-Dihydromegestrol;(6S,8R,9S,10R,13S,14S,17R)-17-acetyl-17-hydroxy-6,10,13-trimethyl-2,6,7,8,9,11,12,14,15,16-decahydro-1H-cyclopenta[a]phenanthren-3-one;(6S,8R,9S,10R,13S,14S,17R)-17-ethanoyl-17-hydroxy-6,10,13-trimethyl-2,6,7,8,9,11,12,14,15,16-decahydro-1H-cyclopenta[a]phenanthren-3-one;4-Pregnen-6a-Methyl-17-ol-3,20-dione-d7;(6S,8R,9S,10R,13S,14S,17R)-17-acetyl-17-hydroxy-6,10,13-triMethyl-6,7,8,9,10,11,12,13,14,15,16,17-dodecahydro-1H-cyclopenta[a]phenanthren-3(2H)-one;Medroxyprogesteroneacetate-13C2-d3;6alpha-Methyl-17alpha-hydroxypregn-4-ene-3,20-dione
Detection method	Colorimetric
Sample type	Plasma, Serum, Urine
Assay type	Quantitative (Competitive)
Range	1.56 - 100 ng/mL
Sensitivity	0.5 ng/mL
Precision	CV
Recovery	80-120%
Shipping	2-8 °C
Stability and Storage	The stability of ELISA kit is determined by the loss rate of activity. The loss rate of this kit is less than 10% prior to the expiration date under appropriate storage condition